



## Calibration Certificate

No. 06931234/001

Model S0141  
Serial Number 06931234  
Number of pages 2

**Process of Calibration** Direct comparison

**Conditions of Measurement** Temperature 23 °C ± 3 °C  
Relative Humidity 40 % ± 20 %

### Calibration Standards (valid to)

Multimeter Datron1081, PE101, 19145, (13.1.2007)

Simulator Pt1000/3850ppm, Pt005, (9.5.2007)

All standards are traceable to ČMI (Czech Metrology Institute) or in terms of relative humidity to DKD Testo (Deutscher Kalibrierdienst).

### Measuring Results

#### Channel 1-Temperature Extern

Standard	Device	Uncertainty	Used Standard	Note
-90.0 °C	-89.9 °C	0.16 °C	Pt005	Pt1000TG8/E 1745/0306
0.0 °C	0.0 °C	0.11 °C	Pt005	
22.40 °C	22.3 °C	0.18 °C	PE101	
80.0 °C	80.0 °C	0.11 °C	Pt005	
150.0 °C	150.0 °C	0.11 °C	Pt005	
260.0 °C	260.0 °C	0.11 °C	Pt005	

#### Channel 2-Temperature Extern

Standard	Device	Uncertainty	Used Standard	Note
-90.0 °C	-89.9 °C	0.16 °C	Pt005	Pt1000TG8/E 2057/0306
0.0 °C	0.0 °C	0.11 °C	Pt005	
22.40 °C	22.4 °C	0.18 °C	PE101	
80.0 °C	80.0 °C	0.11 °C	Pt005	
150.0 °C	150.0 °C	0.11 °C	Pt005	
260.0 °C	259.9 °C	0.11 °C	Pt005	

#### Channel 3-Temperature Extern

Standard	Device	Uncertainty	Used Standard	Note
-90.0 °C	-89.9 °C	0.16 °C	Pt005	Pt1000TG8/E 2048/0306
0.0 °C	0.0 °C	0.11 °C	Pt005	
22.40 °C	22.5 °C	0.18 °C	PE101	
80.0 °C	80.0 °C	0.11 °C	Pt005	
150.0 °C	150.0 °C	0.11 °C	Pt005	

260.0 °C	260.0 °C	0.11 °C	Pt005	
----------	----------	---------	-------	--

*Channel 4-Temperature Extern*

Standard	Device	Uncertainty	Used Standard	Note
-90.0 °C	-89.9 °C	0.16 °C	Pt005	Pt1000TG8/E 2062/0306
0.0 °C	-0.0 °C	0.11 °C	Pt005	
22.40 °C	22.4 °C	0.18 °C	PE101	
80.0 °C	80.0 °C	0.11 °C	Pt005	
150.0 °C	150.0 °C	0.11 °C	Pt005	
260.0 °C	260.0 °C	0.11 °C	Pt005	

**Measuring Uncertainty**

The expanded uncertainty of measurement corresponding to the measurement results is started as the standard uncertainty of measurement multiplied by the coverage factor  $k=2$ . Usually the values is located in the corresponding interval with probability of approximately 95%. This was determined in accordance with EA4/02.

Date of Calibration

Calibrated By

Has Approved

10.5.2006

.....

Jan Pavlica

.....

Milan Jurajda